

# Bordetella Pertussis & Bordetella Parapertussis Real Time PCR Kit

User Manual

For In Vitro Diagnostic Use Only

REF RD-0266-02

For use with ABI Prism® 7000/7300/7500/7900/Step One Plus; iCycler iQ™ 4/iQ™ 5;  
Smart Cycler II; Bio-Rad CFX 96; Rotor Gene™ 6000; Mx3000P/3005P; MJ-Option2/Chromo4;  
Light Cycler™ 480 Instrument

EC REF Obelis S.A.  
Boulevard Général Wahis 53  
1030 Brussels, BELGIUM  
Tel: +(32) 2.732.59.54  
Fax: +(32) 2.732.60.03  
E-Mail: mail@obelis.net

Shanghai ZJ Bio-Tech Co., Ltd.  
www.liferiver.com.cn Tel: +86-21-34680596  
trade@liferiver.com.cn Fax: +86-21-34680595  
2<sup>nd</sup> floor, No. 15 Building, No. 188 Xinjunhuan Road,  
PuJiang Hi-tech Park, Shanghai, China

## 1. Intended Use

Bordetella pertussis & Bordetella parapertussis real time PCR Kit is used for the distinguishing of bordetella pertussis & bordetella parapertussis by real time PCR systems in samples like nasal and pharyngeal secretions, sputum, and etc.

## 2. Principle of Real-Time PCR

The principle of the real-time detection is based on the fluorogenic 5' nuclease assay. During the PCR reaction, the DNA polymerase cleaves the probe at the 5' end and separates the reporter dye from the quencher dye only when the probe hybridizes to the target DNA. This cleavage results in the fluorescent signal generated by the cleaved reporter dye, which is monitored real-time by the PCR detection system. The PCR cycle at which an increase in the fluorescence signal is detected initially (Ct) is proportional to the amount of the specific PCR product. Monitoring the fluorescence intensities during Real Time allows the detection of the accumulating product without having to re-open the reaction tube after the amplification.

## 3. Product Description

Bordetella pertussis and Bordetella parapertussis are closely related species. Both are responsible for outbreaks of whooping cough in humans and produce similar virulence factors, with the exception of pertussis toxin, specific to B. pertussis. B. pertussis causes severe coughing spells, with a characteristic "whoop" made as the affected person struggles to breathe through narrowed airway passages between coughing spasms. B. pertussis is a small gram-negative aerobic coccobacillus that colonizes the cilia of the nose and throat of infected humans. Pertussis caused by B. parapertussis manifests with similar symptoms to B. pertussis-derived disease but in general tends to be less severe.

B. pertussis & B. parapertussis real time PCR kit contains a specific ready-to-use system for the detection of B. pertussis and B. parapertussis by polymerase chain reaction (PCR) in the real-time PCR system. The master contains reagents and enzymes for the specific amplification of the B. pertussis and B. parapertussis DNA. Fluorescence is emitted and measured by the real time systems' optical unit during PCR. The detection of amplified B. pertussis DNA fragment is performed in fluorimeter channel FAM. The detection of amplified B. parapertussis DNA fragment is performed in fluorimeter channel HEX. DNA extraction buffer is available in the kit. In addition, the kit contains a system to identify possible PCR inhibition by measuring the Cal Red 610/ROX/TEXAS RED fluorescence of the internal control (IC). An external positive control (1×10<sup>7</sup> copies/ml) contained, allow the determination of the gene load. For further information, please refer to section 9.3 Quantitation.

## 4. Kit Contents

Ref.	Type of Reagent	Presentation 25rxns
1	DNA Extraction Buffer	2 vials, 1.5ml
2	B. (para)pertussis Reaction Mix	1 vial, 950µl
3	PCR Enzyme Mix	1 vial, 12µl
4	Molecular Grade Water	1 vial, 400µl
5	Internal Control(IC)	1 vial, 30µl
6	B. (para)pertussis Positive Control (1×10 <sup>7</sup> copies/ml)	1 vial, 30µl

Analysis sensitivity: 1×10<sup>3</sup> copies/ml; LOQ: 2×10<sup>3</sup>~1×10<sup>8</sup> copies/ml

Note: Analysis sensitivity depends on the sample volume, elution volume, nucleic acid extraction methods and other factors. If you use the DNA extraction buffer in the kit, the analysis sensitivity is the same as it declares. However, when the sample volume is dozens or even hundreds of times greater than elution volume by some concentrating method, it can be much higher.

## 5. Storage

- All reagents should be stored at -20°C. Storage at +4°C is not recommended.
- All reagents can be used until the expiration date indicated on the kit label.
- Repeated thawing and freezing (>3x) should be avoided, as this may reduce the sensitivity of the assay.
- Cool all reagents during the working steps.
- Reaction Mix should be stored in the dark.

## 6. Additionally Required Materials and Devices

- Biological cabinet
- Trypsin digestive Solution
- Real time PCR reaction tubes/plates
- Pipets (0.5 µl – 1000 µl)
- Sterile microtubes
- Biohazard waste container
- Desktop microcentrifuge for "ependorf" type tubes (RCF max. 16,000 x g)
- Real time PCR system
- Vortex mixer
- Cryo-container
- Sterile filter tips for micro pipets
- Disposable gloves, powderless
- Refrigerator and Freezer
- Tube racks

## 7. Warnings and Precaution

- Carefully read this instruction before starting the procedure.
- For in vitro diagnostic use only.
- This assay needs to be carried out by skilled personnel.
- Clinical samples should be regarded as potentially infectious materials and should be prepared in a laminar flow hood.
- This assay needs to be run according to Good Laboratory Practice.
- Do not use the kit after its expiration date.
- Avoid repeated thawing and freezing of the reagents, this may reduce the sensitivity of the test.
- Once the reagents have been thawed, vortex and centrifuge briefly the tubes before use.
- Prepare quickly the Reaction mix on ice or in the cooling block.
- Set up two separate working areas: 1) Isolation of the RNA/ DNA and 2) Amplification/ detection of amplification products.
- Pipets, vials and other working materials should not circulate among working units.
- Use always sterile pipette tips with filters.
- Wear separate coats and gloves in each area.
- Avoid aerosols

## 8. Sample Collection, Storage and transport

- Collect samples in sterile tubes;
- Specimens can be extracted immediately or frozen at -20°C to -80°C.
- Transportation of clinical specimens must comply with local regulations for the transport of etiologic agents

## 9. Procedure

### 9.1 DNA-Extraction

DNA extraction buffer is contained in the kit. Please thaw the buffer thoroughly and spin down briefly in the centrifuge before use.

#### 9.1.1 Sputum samples

##### 1) Trypsin digestive Solution preparation

Add 10g trypsin to 200ml purified water and mix thoroughly. Adjust PH value to 8.0 with 2% NaOH solution. Add 2mL CaCl<sub>2</sub> (25mmol/L), mix thoroughly and store at 4°C. Please incubate at 37°C for 10 minutes before use.

2) Estimate the volume of the sputum and add nartes aequales of the Trypsin digestive Solution then

vortex vigorously. Set at room temperature for 30 minutes. Transfer 0.5ml mixture to a new tube. Centrifuge the tube at 13000rpm for 5 minutes, carefully remove and discard supernatant from the tube without disturbing the pellet.

3) Add 1.0ml normal saline. Resuspend the pellet with vortex vigorously. Centrifuge at 13000rpm for 5 minutes. Carefully remove and discard supernatant from the tube without disturbing the pellet.

4) Repeat step 3)

5) Add 100µl DNA extraction buffer, close the tube then suspend the pellet with vortex vigorously. Spin down briefly in a table centrifuge.

6) Incubation the tube for 10 minutes at 100°C.

7) Centrifuge the tube at 13000rpm for 10 minutes. The supernatant contains DNA extracted and is used for PCR template.

### 9.1.2 Nasal and pharyngeal secretions samples

1) Take 1ml sample in a tube, centrifuge the tube at 13000rpm for 2min, and remove the supernatant and keep the pellet.

2) Add 100µl DNA extraction buffer to the pellet, close the tube then vortex for 10 seconds. Spin down briefly in a table centrifuge.

3) Incubate the tube for 10 minutes at 100°C.

4) Centrifuge the tube at 13000rpm for 10 minutes. The supernatant contains the DNA extracted and can be used for PCR template.

### Attention:

A. During the incubation, make sure the tube is not open, as the vapor will volatilize into the air and may cause contamination in case the sample is positive.

B. The extraction sample should be used in 3 hours or store at -20°C for one month.

C. DNA extraction kits are available from various manufacturers. You can also use your own extraction systems or the commercial kit depending on the yield. For DNA extraction, please comply with the manufacturer's instructions.

## 9.2 Internal Control

It is necessary to add internal control (IC) in the reaction mix. Internal Control (IC) allows the user to determine and control the possibility of PCR inhibition.

Add the internal control (IC) 1µl/rxn and the result will be shown in the Cal Red 610/ROX/TEXAS RED.


## 9.3 Quantitation

The kit can be used for quantitative or qualitative real-time RT-PCR. A positive control defined as 1×10<sup>7</sup> copies/ml is supplied in the kit.

For performance of quantitative real-time PCR, Standard dilutions must prepare first as follows. Molecular Grade Water is used for dilution.


Dilution is not needed for qualitative real-time PCR detection.

Take positive control (1×10<sup>7</sup> copies/ml) as the starting high standard in the first tube. Respectively pipette 36µl Molecular Grade Water into next three tubes. Do three dilutions as the following figures:



## 9.4 PCR Protocol

The Master Mix volume for each reaction should be pipetted as follows:



- The volumes of Reaction Mix and Enzyme Mix per reaction multiply with the number of samples, which includes the number of controls, standards, and sample prepared. Molecular Grade Water is used as the negative control. For reasons of unprecise pipetting, always add an extra virtual sample. Mix completely then spin down briefly in a centrifuge.
- Pipet 36µl (22.5µl for Smart Cycler II) Master Mix with micropipets of sterile filter tips to each Real time PCR reaction plate/tubes. Separately add 4µl (2.5µl for Smart Cycler II) DNA sample, positive and negative controls to different reaction plate/tubes. Immediately close the plate/tubes to avoid contamination.
- Spin down briefly in order to collect the Master Mix in the bottom of the reaction tubes.
- Perform the following protocol in the instrument:

37°C for 2min	1cycle
94°C for 2min	1cycle
93°C for 15sec, 60°C for 1min	40cycles
(Fluorescence measured at 60°C)	

Selection of fluorescence channels	
FAM	B. pertussis
HEX/VIC/JOE	B. parapertussis
Cal Red 610/ROX/TEXAS RED	IC

- If you use ABI Prism® system, please choose "none" as passive reference and quencher.

10. Threshold setting: just above the maximum level of molecular grade water.

11. Calibration for quantitative detection: Input each concentration of standard controls at the end of run, and a standard curve will be automatically formed.

12. Quality control: Negative control, positive control, internal control and QS curve must be performed correctly, otherwise the sample results is invalid.

Control	Channel			Ct value	
	FAM	HEX/VIC/JOE	Cal Red 610	UNDET	UNDET
Molecular Grade Water	UNDET	UNDET	25~35		
Positive Control(qualitative assay)	≤35	≤35	—		
QS (quantitative detection)	Correlation coefficient of QS curve≤-0.98				

## 13. Data Analysis and Interpretation :The following sample results are possible:

Ct value			Result Analysis
FAM	HEX	Cal Red 610	
1# UNDET	UNDET	25~35	Below the detection limit or negative
2# ≤38	UNDET	—	B. pertussis Positive; and the software displays the quantitative value
3# UNDET	≤38	—	B. parapertussis Positive; and the software displays the quantitative value
4# 38~40	25~35	—	Re-test; If it is still 38~40, report as 1#

For further questions or problems, please contact our technical support at [trade@liferiver.com.cn](mailto:trade@liferiver.com.cn)